

LDH (P-L) KIT

(DGKC Method)

(For invitro diagnostic use only)



INTENDED USE


LDH (P - L) Kit is used for the determination of Lactate Dehydrogenase Activity in serum.

SUMMARY

LDH is found in many body tissues particularly heart, liver, skeletal muscle, kidney and RBC'S. LDH is found in the form of isoenzymes based on their electrophoretic mobility with each isoenzyme being primarily from different organs. Increased levels are found in myocardial infarction, pulmonary diseases, hepatic diseases, hemolytic anemias, renal diseases and muscular dystrophy.

PRINCIPLE

Lactate dehydrogenase catalyzes the reduction of pyruvate with NADH to form NAD. The rate of oxidation of NADH to NAD is measured as a decrease in absorbance, which is proportional to the LDH Activity in the sample.



EXPECTED VALUES

Serum : 230 - 460 U/L at 37°C

It is recommended that each laboratory establish its own normal range representing its patient population.

PRESENTATION

| REF | 1102160025 | 1102160275 | 1102160075 |
|-----------|------------|------------|------------|
| Pack Size | 25 ml | 2 x 75 ml | 75 ml |

| | | | |
|--------------------|-------|-----------|-------|
| L1 Buffer Reagent | 20 ml | 2 x 60 ml | 60 ml |
| L2 Starter Reagent | 5 ml | 2 x 15 ml | 15 ml |

COMPOSITION

Tris Buffer 100mM; Pyruvate 1.2mM; NaCl 200mM; NADH 0.18mM; Non Reactive Stabilizers and Preservatives.

STORAGE / STABILITY

Contents are stable at 2-8°C till the expiry mentioned on the labels.

REAGENT PREPARATION

Reagents are ready to use.

Working reagent: For sample start assays a single reagent is required. Pour the contents of 1 bottle of L2 (Starter Reagent) into 1 bottle of L1 (Buffer Reagent). This working reagent is stable for at least 1 week when stored at 2-8°C.

Alternatively for flexibility as much of working reagent may be made as and when desired by mixing together 4 parts of L1 (Buffer Reagent) and 1 part of L2 (Starter Reagent). Alternatively 0.8 ml of L1 and 0.2 ml of L2 may also be used instead of 1 ml of the working reagent directly during the assay.

SAMPLE MATERIAL

Serum. Free from hemolysis. Total LDH is reported to be stable in serum for 1-3 days at 2-8°C. **Freezing inactivates the liver isoenzyme.**

SAMPLE WASTE AND DISPOSAL

Do not reuse the reagent containers, bottles, caps or plugs due to the risks of contamination and the potential to compromise reagent performance.

This product requires the handling of human specimens. It is recommended that all human sourced material are considered potentially hazardous and are handled in accordance with the OSHA standard on blood borne pathogens.

Appropriate biosafety practices should be used for materials that contain or are suspected of containing infectious agents.

Handle specimens, solid and liquid waste and test components in accordance with local regulations and NCCLS guidelines M29, or other published biohazard safety guidelines.

MATERIALS REQUIRED BUT NOT PROVIDED

Photometer analyzer with standard thermostatic cuvette holder, micropipette and appropriate laboratory equipment.

PROCEDURE

| | | |
|---------------------|---|--------------------|
| Wavelength / filter | : | 340 nm |
| Temperature | : | 37°C / 30°C / 25°C |
| Light path | : | 1 cm |

Substrate Start Assay:

Pipette into a clean dry test tube labelled as Test (T):

| Addition Sequence | (T) 25°C / 30°C | (T) 37°C |
|--|-----------------|----------|
| Buffer Reagent (L1) | 0.8 ml | 0.8 ml |
| Sample | 0.05 ml | 0.02 ml |
| Incubate at the assay temperature for 1 min. and add | | |
| Starter Reagent (L2) | 0.2 ml | 0.2 ml |

Mix well and read the initial absorbance A_0 after 1 min. and repeat the absorbance reading after every 1, 2, & 3 mins. Calculate the mean absorbance change per min. ($\Delta A / \text{min}$).

Sample Start Assay:

Pipette into a clean dry test tube labelled as Test (T):

| Addition Sequence | (T) 25°C / 30°C | (T) 37°C |
|--|-----------------|----------|
| Working Reagent | 1.0 ml | 1.0 ml |
| Incubate at the assay temperature for 1 min. and add | | |
| Sample | 0.05 ml | 0.02 ml |

Mix well and read the initial absorbance A_0 after 1 min. and repeat the absorbance reading after every 1, 2, & 3 mins. Calculate the mean absorbance change per min. ($\Delta A / \text{min}$).

CALCULATIONS

| | | |
|---------------------------------|-------------|---------------------------------------|
| Substrate / Sample start | | |
| LDH Activity in U/L | 25°C / 30°C | = $\Delta A / \text{min} \times 3333$ |
| | 37°C | = $\Delta A / \text{min} \times 8095$ |

QUALITY CONTROL

The following process is recommended for QC during the assay of LDH (P-L). *Define and establish acceptable range for your laboratory.

- Two levels of control (Normal and Abnormal) are to be run on a daily basis.
- If QC results fall outside acceptance criteria, recalibration may be necessary.
- Review QC results and run acceptance criteria following a change of reagent lot.

SPECIFIC PERFORMANCE CHARACTERISTICS

Linearity:

The procedure is linear upto 2000 U/L at 37°C. If the absorbance change ($\Delta A / \text{min}$) exceeds 0.250, use only the value of the first 2 mins. to calculate the result or dilute the sample 1+9 with normal saline (NaCl 0.9%) and repeat the assay (Results x 10).

Limit of detection:

The limit of detection for LDH (P-L) is 8 U/L.

Interferences:

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

Precision:

Precision studies were performed with two controls using NCCLS protocol EP5-A. The results of the precision studies are shown below:

| Sample | Within-run | | Between-run | | Total | |
|-----------|------------|------|-------------|------|--------|------|
| | Mean | CV% | Mean | CV% | Mean | CV% |
| Control 1 | 331.86 | 1.80 | 332.54 | 1.71 | 664.4 | 3.51 |
| Control 2 | 524.01 | 1.22 | 527.69 | 0.86 | 1051.7 | 2.08 |

Method comparison:

Comparative studies were done to compare our reagent with another commercial LDH (P-L) Assay. No significant differences were observed. Details of the comparative studies are available on request.

TEMPERATURE CONVERSION FACTORS

| Assay Temperature | Desired Reporting Temperature | | |
|-------------------|-------------------------------|------|------|
| | 25°C | 30°C | 37°C |
| 25°C | 1.00 | 1.33 | 1.92 |
| 30°C | 0.75 | 1.00 | 1.44 |
| 37°C | 0.52 | 0.70 | 1.00 |

NOTE

In vitro diagnostic reagent for laboratory and professional use only Not for medicinal use. The reagent contain sodium azide 0.1% as preservative. Avoid contact with skin and mucosa. On disposal flush with large quantities of water. Only clean and dry glassware must be used. Samples having a high activity show a very low initial absorbance as most of the NADH is consumed prior to the start of measurement. If this is suspected then dilute the sample and repeat the assay.

The working reagent or the combined reagent should have an absorbance above 1.000 against distilled water at 340 nm. Discard the reagent if the absorbance is below 1.000.

RBC's have a very high LDH content and hence hemolysed samples should not be used. The reagent may be used in several automated analyzers. Instructions are available on request. Do not use turbid, deteriorated or leaking reagents.

REFERENCES

- Recommendations for the measurement of LDH in human serum at 30°C. Ann. Biol. Chem. (1982) 40: 87.
- Weishaar H. D. et. al. (1975) Med. Welt 26:387 Calculated Values of Section. Please Note "Temperature Conversion Factors".
- Data on file: Coral Clinical Systems.

| System Parameters | | | |
|-------------------|-------------------|--------------|--------------|
| Reaction | : U.V. Kinetic | Interval | : 60 Sec. |
| Wavelength | : 340 nm | Sample Vol. | : 0.02 ml |
| Zero Setting | : Distilled Water | Reagent Vol. | : 1.00 ml |
| Incub. Temp. | : 37°C | Standard | : --- |
| Incub. Time | : --- | Factor | : 8095 |
| Delay Time | : 60 Sec. | React. Slope | : Decreasing |
| Read Time | : 180 Sec. | Linearity | : 2000 U/L |
| No. of read. | : 4 | Units | : U/L |

| | | | | |
|-----------------------------------|------------------------------|--|---------------------------|--------------------------------------|
| Store at 2-8°C | Manufacturer | IVD In vitro Diagnostic Medical Device | L1 Buffer Reagent | Mod. IFCC Mod. IFCC Method |
| Use by (Last day of stated month) | Consult Instructions for use | LOT Batch Number | L2 Starter Reagent | This way up |
| Date of Manufacture | REF Catalogue Number | EC REP Authorised Representative in the European Community | | |



Manufactured by:

Coral Clinical Systems

A Division of Tulip Diagnostics (P) Ltd.

PLOT NO. M-46, PHASE III B, VERNA IND. EST., VERNA, GOA-403 722, INDIA.

REGD. OFFICE : GITANJALI, TULIP BLOCK, DR. ANTONIO DO REGO BAGH, ALTO SANTACRUZ, BAMBOLIM COMPLEX P.O., GOA-403 202, INDIA.

EC REP

CMC Medical Devices & Drugs S.L.,
C/ Horacio Lengo No. 18, CP 29006, Malaga, Spain.

LDH110200/VER-03